



**National Accreditation Board for  
Testing and Calibration Laboratories**  
(A Constituent Board of Quality Council of India)



## **CERTIFICATE OF ACCREDITATION**

**ANALYTICAL TECHNOLOGY LABORATORY, SRI  
GOMUKI TEX CHEM PVT. LTD.**

has been assessed and accredited in accordance with the standard

**ISO/IEC 17025:2005**

"General Requirements for the Competence of Testing & Calibration Laboratories"

for its facilities at

1/93, Mylampatti P.O., Coimbatore, Tamil Nadu

in the field of

**TESTING**

**Certificate Number** TC-7341

**Issue Date** 06/06/2018

**Valid Until** 05/06/2020

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website [www.nabl-india.org](http://www.nabl-india.org))

Signed for and on behalf of NABL



89076970100030001455

**Anil Relia**  
Chief Executive Officer



# National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



## SCOPE OF ACCREDITATION

**Laboratory** Analytical Technology Laboratory, Sri Gomuki Tex Chem Pvt. Ltd.,  
1/93, Mylampatti P.O., Coimbatore, Tamil Nadu

**Accreditation Standard** ISO/IEC 17025: 2005

**Certificate Number** TC-7341

Page 1 of 1

**Validity** 06.06.2018 to 05.06.2020

Last Amended on --

Sl.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
-----	----------------------------	-------------------------	---	--

### CHEMICAL TESTING

<b>I.</b>	<b>FOOD &amp; AGRICULTURAL PRODUCTS</b>			
<b>1.</b>	<b>Starch and Starch Products</b>	Moisture	IS 4706 (Part 2)	(1 g to 20 g) per 100 g
		pH	IS 4706 (Part 2)	4.00 to 10.00
		Total Ash	IS 4706 (Part 2)	(0.02 g to 25 g) per 100 g
<b>2.</b>	<b>Carboxy Methyl Cellulose</b>	Loss on Drying	IS 5306 (Annexure C)	0.1 g/100g to 25 g/100g
<b>II.</b>	<b>INDUSTRIAL &amp; FINE CHEMICALS</b>			
<b>1.</b>	<b>PolyVinyl alcohol</b>	Volatile matter	IS 11840 (Appendix A-5)	0.1 g/100g to 10 g/100g
		pH	IS 11840 (Appendix A-8)	4.00 to 10.00
		Viscosity	IS 11840 (Appendix A-9)	5 cP to 100 cP
<b>III.</b>	<b>WATER</b>			
<b>1.</b>	<b>Industrial Water</b>	Turbidity	IS 3025 (Part 10)	0.5 NTU to 100 NTU
		Odour	IS 3025 (Part 05)	Qualitative (Agreeable / Disagreeable)
		pH	IS 3025 (Part 11)	4 to 10
		Specific Conductance	IS 3025 (Part 14)	0.1 $\mu$ S/cm to 1413 $\mu$ S/cm
		Total Residue	IS 3025 (Part 15)	1 mg/L to 1500 mg/L
		Non-Filterable Residue	IS 3025 (Part 17)	4 mg/L to 500 mg/L
	Filterable Residue	IS 3025 (Part 16)	1 mg/L to 1500 mg/L	

Naveen Jangra  
Convenor

Alok Jain  
Program Director